









EC declaration of conformity

We, MSP Production B.V.
De Sonman 7
5066 GJ Moergestel
The Netherlands

Declare that: Electrically adjustable Shower/Toilet Seat
Brand Name Noa

Type: Noa Standard 135, Noa Standard 190,
Noa Low 135, Noa Low 190

With serial numbers MPNOA***** (* has a value between 0-9)

- Is compliant with Directive MDR EU 2017/745 (Medical Devices Regulation) and is classified as Class 1
- The Noa has been safety tested in accordance with:
 -  EN 10535
 -  EN 12182
 -  EN 60601-1
 -  Bio compatibility ISO 10993
- The Noa has been EMC tested and meets all requirements
 -  Emission EN 60601-1-2 (2007) + AC (2010)
 -  Emission EN 61000-3-2 (2006) + A1 (2009) + A2 (2009)
 -  Emission EN 61000-3-3 (2008)
 -  Immunity EN 60601-1-2 (2007) + AC (2010)
- The internal power source/battery has been safety tested according to IEC60601-1, ANSI/AAMI ES60601-1, CAN/CSA-22.2 No 60601-1, IEC62133, UL2054, UN38.3 (needed for transport of lithium batteries)

MSP-Production B.V. is not aware of any electromagnetic disturbances caused by interactions between the Noa and other electrical products.

Technical file compiled by L.J.M.G. Dhanpat.

The Netherlands, Moergestel, 01-10-2022

L.J.M.G. Dhanpat, Managing Director

MSP Production B.V.

